



neolab Migge GmbH  
Rischerstr. 7-9  
69123 Heidelberg  
Deutschland  
+49 (0)6221 /  
8442-44  
<https://www.neolab.de>  
e


Umsatzsteuer-  
Identifikationsnummer  
:  
DE 143 450 657



## Brand Dispensette® S, Analog, DE-M 5 - 50 ml, with back-dispensing valve

**€534.90**  
**plus VAT &**  
**Shipping**

### Product Images



## Description

---

The Dispensette® S bottle-top dispenser is the number 1 for dispensing directly from the supply bottle. It is equipped with everything that makes dispensing liquids safer and pleasantly easy. Thanks to the special operating principle of the piston, the Dispensette® S is smooth and effortless to operate during liquid aspiration and dispensing. The volume can be quickly and easily adjusted and securely fixed. The Dispensette® S bottle-top dispenser is proven in demanding continuous use and in the use of aggressive media. Dispensette® S (color code red):

Ideal for dispensing aggressive reagents, e.g. concentrated alkalis and acids such as  $\text{HPO}_4$ ,  $\text{H}_2\text{SO}_4$  (except e.g.  $\text{HCl}$ ,  $\text{HNO}_3$  and  $\text{HF}$ ), salt solutions and a variety of organic solvents. Analog version:

Analog volume display, quickly adjustable dosing volume, securely fixed by internal toothed bar. Easy adjustment with adjustment key.

## Additional Information

No.	BN-1331
Article - USP	Minimum operating forces, maximum adjustment comfort. Proven in demanding continuous operation and with aggressive media. The number 1 for dispensing directly from the bottle.
Manufacturer (Brand)	Brand
VGKL number	104946161
EAN	4058072207021
Gross weight	0.87
HSNumber	84131900
Storage temperature	Room temperature
Transport temperature	Room temperature
Net weight	0.65600 kg
Volume MAX	50 ml
Volume MIN	5 ml
Country of origin	Germany
Packing width	0.20400 m
Packing height	0.09000 m
Packing depth	0.28500 m
Packaging volume	5232.60000000 ccm
Customs tariff number	84131900
Dispenser properties	with back metering valve
Division	1 ml
Measurement deviation systematic in %	0.5 %
Measurement deviation systematic in µl	250 µl
Measurement deviation random in µl	50 µl
Youtube URL	<a href="https://youtu.be/GzpO7WPlixw">https://youtu.be/GzpO7WPlixw</a>
Youtube URL 2	<a href="https://youtu.be/VKPH3Y6_z4Y">https://youtu.be/VKPH3Y6_z4Y</a>
for thread/DM	GL 32-33. GL 38. S 40

