



neolab Migge GmbH
Rischerstr. 7-9
69123 Heidelberg
Deutschland
+49 (0)6221 /
8442-44
<https://www.neolab.de>
e

Umsatzsteuer-
Identifikationsnummer
:
DE 143 450 657



Eppendorf Reference® 2, 2 – 20 µL, light gray

€376.00
plus VAT &
Shipping

Product Images



Description

The Reference®2 microliter pipette offers exceptional precision.

With its ergonomic design, variable design for volumes and long service life, it is a reliable partner for demanding work in diagnostics, research and industry.

- Simple and visual volume adjustment with volume fixation
- Volume range from 0.1 µl to 10 ml
- Fully autoclavable
- Single button operation
- Reproducible tip fit on the pipette cone

Each variable Reference up to 1000 µl contains an epT.I.P.S. box with tips of the corresponding size.

Pipettes with a larger final volume (2.5/5.0/10 ml) are supplied with an epT.I.P.S.® sample bag.

Additional Information

| | |
|--|--|
| No. | VB-1491 |
| Manufacturer (Brand) | Eppendorf |
| old neoLab article no. | E-0432 |
| EAN | 4058072416027 |
| Transport temperature | Room temperature |
| autoclavable at 121°C | Yes |
| Color | White |
| Volume MAX | 20 µl |
| Volume MIN | 2 µl |
| sterile | No |
| Operation via | Knobs/Buttons |
| Microliter pipettes features | with certification with GLP functions with color code ergonomically shaped |
| Number of channels | 1 |
| Energy supply | without power supply |
| Color code | light gray |
| Functionality | mechanical |
| Model Microliter pipette/ Dispenser | Reference 2 2 - 20µl |
| Division | 0.02 µl |
| Type Microliter pipette (design) | Piston stroke pipette air cushion pipette |
| Manufacturer | Eppendorf |
| Measurement deviation systematic in % | 0.8 % |
| Measurement deviation systematic in µl | 0.16 µl |
| Measurement deviation random in % | 0.3 % |
| Measurement deviation random in µl | 0.06 µl |
| Type pipette | Microliter pipettes |
| Type certificate | Declaration of conformity (standardization) |

